

NIOSH N95 APPROVED
n h a l e x h a l e
 r e s p i r a t o r s
 Model NX952 Box of 20



change the design

n h a l e x h a l e
 r e s p i r a t o r s

Instructions for the NX952 N95 Maintenance Free Respirator

Fitting Instructions

The respirator user must be fit tested according to OSHA regulation 29 CFR 1910.134.



1. Hold the respirator in one hand with the nosepiece at your fingertips, allowing the headbands to hang below your hand.



2. Press the respirator against your face with the nosepiece on the bridge of your nose.



3. Place the top band high on the back of your head. Move the bottom band over your head and position it below your ears.



4. Using both hands, mould the nosepiece to the shape of your nose.



5. Test the fit. Cup both hands over the respirator and exhale vigorously. If air flows around your nose, tighten the nosepiece. If air leaks around the edges, reposition the bands for better fit.

If you CANNOT achieve a proper fit, DO NOT enter a contaminated area. See your supervisor for proper fitting respiratory protection.

PARTICULATE RESPIRATOR
RESPIRADOR PARA LIJAR PINTURA

RESPIRADOR PARA LIJAR PINTURA



MODEL NX952

- Cushioned Nose Pad
- Almohadilla para la nariz acolchada
- Two Latex-Free Straps
- Dos correas sin látex
- Adjustable Aluminum Nosepiece
- Puente nasal de aluminio ajustable
- NIOSH Approved
- Aprobados por NIOSH

N95 APPROVED
filters at least 95% of airborne particles
20 RESPIRATORS

Manufacture Date
 Expiration Date
 Lot No.

APPLICATIONS

USE AGAINST:

NIOSH approved for protection against particulate aerosols free of oils. Cordova Safety Products recommends this respirator for use against particulates such as flour, iron coal and minerals.

DO NOT USE AGAINST:

1. Vapors, paint spray, oil-based aerosols, asbestos and sand blasting.
2. Any contaminant whose concentration is unknown, or immediately dangerous to life.
3. Concentrations of Dust and Mists which exceed the maximum use concentration or 10 times contaminant's permissible exposure limit established by OSHA.

WARNING:

1. These respirators do not supply oxygen. Use only in well ventilated areas containing sufficient oxygen to support life.
2. Leave contaminated area immediately if the respirator becomes damaged and proper fit cannot be obtained or breathing becomes difficult. Dispose and change the respirator.
3. The respirator has been tested. Small punctures around the staples are normal and do not interfere with the respirator compliance with Part 84 approval requirements. However, it is subject to inspect prior to each use to assure there are no holes in the breathing zone other than the punctures around staples and no damage has occurred. Enlarged holes resulting from ripped or torn filter material around staple punctures are considered damage.

IMPORTANT LIMITATIONS:

1. Before occupational use of this respirator a written respiratory protection program must be implemented meeting all the local government requirements. In the United States employers must comply with OSHA 29 CFR 1910.134 which includes medical evaluation, training and fit testing.
2. All instructions and warnings have to be followed. Failure to do so will reduce the efficiency of respirator and cause sickness or death.
3. Training on the proper use of respirator must be provided by the employer in accordance with the safety and health standard.
4. Do not alter or modify the respirator.
5. Not suitable for persons with beard or other facial hair which may reduce the efficiency of respirator.
6. The use and reuse of this respirator is subject to consideration of hygiene damage and increased breathing resistance. Service time should not be extended beyond 8 total hours of use.
7. Store the respirators in cool place. Avoid direct exposure of sunlight during storage as the filter media in this respirator can be degraded by extended exposure to direct sunlight.

change the design



FILTERING FACEPIECE APPROVAL LABEL

Cordova Safety Products
 4025 Viscount Avenue
 Memphis, TN, 38118
 Phone: 800-458-8763



THIS RESPIRATOR IS APPROVED ONLY IN THE FOLLOWING CONFIGURATION:

TC	Protection	Respirator	Cautions and Limitations*
84A-3888	N95	NX952	ABCIMNOPS

1. Protection

N95- Particulate Filter (95%) filter efficiency level effective against particulate aerosols free of oil. Some use restrictions may apply.

2. Cautions and Limitations

- A. Not for use in atmospheres containing less than 10.5% oxygen.
- B. Not for use in atmospheres immediately dangerous to life or health.
- C. Do not exceed maximum use concentrations established by regulatory standards.
- F. Failure to properly use and maintain this product could result in injury or death.
- M. All approved respirators shall be selected, fitted, used, and maintained in accordance with MSHA, OSHA and other applicable regulations.
- N. Never substitute, modify, add, or omit parts. Use only exact replacement parts in the configuration as specified by the manufacturer.
- O. Refer to User Instructions, and/or maintenance manuals for information on use and maintenance of these respirators.
- P. NIOSH does not evaluate respirators for use as surgical masks.
- Q. Special or critical use instructions and/or specific use limitations apply. Refer to User Instructions before donning.



WARNING

NIOSH APPROVED N95 95% filtration efficiency against solid and liquid aerosols that do not contain oil.

This respirator helps protect against certain particulate contaminants but does not eliminate exposure or the risk of contracting disease or infection. Misuse may result in sickness or death.

Important notice to purchaser:

The following warranty is made in lieu of all other warranties expressed or implied, including but not limited to, the implied warranties of merchantability and fitness for purpose. Manufacturer warrants this product to be free from defects. Manufacturer's only obligation shall be, at our option, to replace any portion proving defective or to refund the purchase price thereof. User shall determine the suitability of the product for his intended use, and assume all risks and liability in connection therewith.

This model NX952 filtering facepiece respirator has been manufactured by Infynity Industrial Co. Ltd. for Cordova Safety Products under NIOSH TC-84A-3888.

Cordova Safety Products
 4025 Viscount Avenue
 Memphis, TN 38118
 800 458 8763
 www.cordova-safety.com